




waag society

institute for art, science and technology

Waag Society – CC-BY-SA 3.0



BioHack Academy Spectrometer Design



Content

- Webcam spectrometer
- CCD Array spectrometer




waag society

institute for art, science and technology

Webcam Spectrometer




BioHack Academy design







BioHack Academy design





BioHack Academy design





Bill of Materials

Amount Description

1 1 Webcam

2 1 3mm MDF

3 1 White LED

4 1 220 Ohm resistor

5 1 7.5 W power supply


6 1 Jack Adapter

7 1 4.7 GB DVD-R (track pitch 0.74 microns) or 1,000 lines/mm grating





Foil

- Angle 45 degrees
 - Grating foil
 - 1000 lines per mm
 - DVD-R 4.7 GB:
 - Track pitch of 0.74 microns
 - 1351 lines per mm
 - DVD-R 3.95 GB:
 - Track pitch of 0.8 microns
 - 1250 lines per mm
 - CD:
 - track pitch 1.6 microns
 - 625 lines per mm






LED Wiring





Slit


- Slit size: 1mm
- Lower is better resolution, but less sensitive
- Distance to slit: 22 cm
 - Focus webcam on this distance






Spectrum

- Day light






Spectral Workbench analysis






Calibration





Calibration



<http://publiclab.org/wiki/spectral-workbench-calibration>




waag society

institute for art, science and technology

CCD Array Spectrometer




BioHack Academy design






BioHack Academy design






BioHack Academy design






Bill of Materials

#	Amount	Description
1		1 White LED
2		1 TSL1406R Photo diode array
3		1 1,000 lines/mm grating
4		1 220 ohm resistor
5		1 Power switch
5		1 7.5 W power supply
6		1 Jack Adapter








CCD array

- 400 nm to 900 nm target wavelength range
- d = grating 0.001 mm / line
- λ = wavelength
- ϕ = diffraction angle
- m = order of diffraction = 1

$$d \times \sin(\theta) = m \times \lambda$$


$$\theta = \sin^{-1}\left(\frac{m \times \lambda}{d}\right)$$

- 400 nm: $\phi = 23.58$ degrees
- 900 nm $\phi = 64.16$ degrees





Processing Software – SpectrumDisplay





Source Code – setup

```
import processing.serial.*;

Serial port;      // The serial port
int exposureTime = 1; // 1 ms
float updateSpeed = 2000; // 1000 ms

int lastSpectrumUpdate=0;

boolean ledState=true;
PFont defaultFont;
PFont titleFont;

String buffer;

int SpectrumSize = 768;
float[] rawSpectrumData, correctedSpectrumData;
float[] darkReadout, whiteReadout;
int spectrumValueIndex=0;
int spectraCount=0;
```

```
void setup() {
    size(800, 400);
    // create a font with the third font available to the system:
    defaultFont = createFont(PFont.list()[0], 14);
    titleFont = createFont(PFont.list()[4], 20);
    textFont(defaultFont);

    // List all the available serial ports:
    printArray(Serial.list());

    String portName = Serial.list()[0];
    port = new Serial(this, portName, 57600);

    correctedSpectrumData = new float[SpectrumSize];
    rawSpectrumData = new float[SpectrumSize];
    darkReadout = new float[SpectrumSize];
    whiteReadout = new float[SpectrumSize];
    for(int i=0;i<SpectrumSize;i++)
        whiteReadout[i]=1.0f;

    lastSpectrumUpdate=millis();

    ledState=true;
    port.write("led 1\n");
}
```



Source Code – spectrum

```
void readSpectrum() {
  println("Updating spectrum");
  port.write("read\n");
}

void drawSpectrum() {
  int xstart=10, ystart=height-30;

  float maxVal=0.0f;
  for (int i=0;i<correctedSpectrumData.length;i++)
    maxVal = max(correctedSpectrumData[i],maxVal);

  float yscale = (height - 160)/maxVal;
  for (int i=1;i<correctedSpectrumData.length;i++) {
    line(i+xstart, ystart - correctedSpectrumData[i-1] * yscale,
        i+xstart+1, ystart - correctedSpectrumData[i] * yscale);
  }

  int indexAtMousePos = max(0, min(SpectrumSize-1, mouseX - xstart));
  text("At mouse: " + correctedSpectrumData[indexAtMousePos], 350, 390);
  text("MaxValue: " + maxVal, 200, 390);
}

/*
Compute correctedSpectrumData based on rawSpectrumData, whiteReadout
*/
void computeSpectrum() {
  for (int i=0;i<SpectrumSize;i++) {
    correctedSpectrumData[i] = (rawSpectrumData[i] - darkReadout[i])
  }
}
```

```
void draw() {
  int time = millis();

  if (lastSpectrumUpdate + updateSpeed < time) {
    lastSpectrumUpdate=time;
    readSpectrum();
  }

  background(255);
  text("U=decrease exposure, I=increase exposure", 10, 55);
  text("Exposure time: " + exposureTime + " ms", 10, 75);
  text("D=set dark measurement", 10, 100);
  text("W=set white measurement (no sample)", 10, 115);

  text("Measurements: " + spectraCount, 10, 390);

  text("LED = " + ( ledState ? "ON" : "OFF" ), 700, 55);
  text("L = Toggle LED", 700, 70);

  textFont(titleFont);
  fill(0, 255, 0);
  text("BioHackAcademy Spectrophotometer", 10, 20);
  textFont(defaultFont);
  fill(0);

  drawSpectrum();
}
```



Source Code – interaction

```
void serialEvent(Serial port) {
  while (port.available () >0) {
    char c = port.readChar();
    if (c == '\n' && buffer.length() > 0) {
      char first=buffer.charAt(0);
      if (first >= '0' && first <= '9') {
        int value = Integer.parseInt(buffer.trim()); //substring(0,buffer.length()-1);
        rawSpectrumData[spectrumValueIndex++] = value/1024.0f;
        if (spectrumValueIndex==SpectrumSize) {
          computeSpectrum();
          spectrumValueIndex=0;
          spectraCount ++;
        }
      } else if(buffer=="start") {
        spectrumValueIndex=0; // align again
      }
      buffer="";
    } else {
      buffer+=c;
    }
  }
}
```

```
void keyPressed() {
  if (key >= 'A' && key <= 'Z')
    key += 'a'-'A'; // make lowercase

  if (key == 'i' || key == 'u') {
    if (key == 'i') exposureTime += 3;
    else exposureTime -= 3;
    exposureTime = max(exposureTime,1);

    port.write("exp " + max(1, (int)exposureTime) + "\n");
  }

  if (key == 'l') {
    ledState=!ledState;
    port.write("led " + ( ledState ? "1" : "0" )+ "\n");
  }


  if (key == ' ') {
    readSpectrum();
  }

  if (key == 'd') {
    for (int i=0;i<SpectrumSize;i++)
      darkReadout[i] = rawSpectrumData[i];
  }

  if (key == 'w') {
    for (int i=0;i<SpectrumSize;i++)
      whiteReadout[i] = rawSpectrumData[i];
  }
}
```




Processing Software – Empty





Coffee





Red food dye



SpectrumDisplay

BioHackAcademy Spectrophotometer

U=decrease exposure, I=increase exposure


Exposure time: 1 ms

D=set dark measurement

W=set white measurement (no sample)

LED = ON


L = Toggle LED



Measurements: 1106

MaxValue: 1.1333333

At mouse: 1.0666667





Green food dye



SpectrumDisplay

BioHackAcademy Spectrophotometer

U=decrease exposure, I=increase exposure


Exposure time: 1 ms

D=set dark measurement

W=set white measurement (no sample)

LED = ON


L = Toggle LED



Measurements: 1113

MaxValue: 1.0

At mouse: 0.93333334






waag society

institute for art, science and technology

Coming up



Practical – Iodine Clock Reaction





Assignment

Hack the spectrometer
into a flow spectrometer;

Add an LCD display;

OD600 measurement.





**some
rights
reserved**

These slides are published by Waag
Society under CC-BY-SA 4.0 license